

# iTaq Universal SYBR® Green Supermix

| Catalog # | Description  |
|-----------|--|
| 1725120   | iTaq Universal SYBR® Green Supermix, 2 ml (2 x 1 ml vials), 200 x 20 µl reactions          |
| 1725121   | iTaq Universal SYBR® Green Supermix, 5 ml (5 x 1 ml vials), 500 x 20 μl reactions          |
| 1725122   | iTaq Universal SYBR® Green Supermix, 10 ml (10 x 1 ml vials), 1,000 x 20 µl reactions      |
| 1725124   | iTaq Universal SYBR® Green Supermix, 25 ml (5 x 5 ml vials), 2,500 x 20 μl reactions       |
| 1725125   | iTaq Universal SYBR® Green Supermix, 50 ml (10 x 5 ml vials), 5,000 x 20 $\mu$ l reactions |

For research purposes only.

#### **Storage and Stability**

Guaranteed for 12 months in a constant temperature freezer at -20°C protected from light. For convenience, this supermix can be stored at 4°C for up to 3 months.

#### **Kit Contents**

iTaq Universal SYBR® Green Supermix is a 2x concentrated, ready-to-use reaction master mix optimized for dye-based quantitative PCR (qPCR) on any real-time PCR instrument (ROX-independent and ROX-dependent). It contains antibody-mediated hot-start iTaq DNA Polymerase, dNTPs, MgCl<sub>2</sub>, SYBR® Green I Dye, enhancers, stabilizers, and a blend of passive reference dyes (including ROX and fluorescein).

## **Instrument Compatibility**

This supermix is compatible with all Bio-Rad and other commercially available real-time PCR systems.

# **Reaction Mix Preparation and Thermal Cycling Protocol**

 Thaw iTaq Universal SYBR® Green Supermix and other frozen reaction components to room temperature. Mix thoroughly, centrifuge briefly to collect solutions at the bottom of tubes, then store on ice protected from light.

- Prepare (on ice or at room temperature) enough reaction mix for all qPCR reactions by adding all required components, except the DNA template, according to the recommendations in Table 1.
- Mix the reaction mix thoroughly to ensure homogeneity and dispense equal aliquots into each qPCR tube or into the wells of a qPCR plate. Good pipetting practice must be employed to ensure assay precision and accuracy.
- 4. Add DNA samples (and nuclease-free H<sub>2</sub>O if needed) to the PCR tubes or wells containing reaction mix (Table 1), seal tubes or wells with flat caps or optically transparent film, and vortex 30 sec or more to ensure thorough mixing of the reaction components. Spin the tubes or plate to remove any air bubbles and collect the reaction mixture in the vessel bottom.
- 5. Program the thermal cycling protocol on a real-time PCR instrument according to Table 2.
- Load the PCR tubes or plate into the real-time PCR instrument and start the PCR run.
- 7. Perform data analysis according to the instrument-specific instructions.

Table 1. Reaction setup.\*

| <u> </u>                                 |                           |                           |  |
|--|---------------------------|---------------------------|--|
| Component                                | Volume/20 μl Reaction, μl | Volume/10 μl Reaction, μl | Final Concentration                            |
| iTaq Universal SYBR® Green Supermix (2x) | 10                        | 5                         | 1x   |
| Forward and reverse primers              | Variable                  | Variable                  | 300-500 nM each primer                         |
| DNA template (add at step 4)             | Variable                  | Variable                  | cDNA: 100 ng–100 fg<br>Genomic DNA: 50 ng–5 pg |
| Nuclease-free H <sub>2</sub> O           | To 20 µl                  | To 10 μl                  | _  |
| Total reaction mix volume                | 20 μΙ                     | 10 µl                     | _  |

<sup>\*</sup> Scale all components proportionally according to sample number and reaction volumes.

Table 2. Thermal cycling protocol.

|  |                   |  | Amplification             |   |        | _   |
|--|-------------------|--|---------------------------|---|--------|---|
| Real-Time PCR System   | Setting/<br>Block | Polymerase Activation and DNA Denaturation at 95°C | Denaturation at 95°C, sec | Annealing/Extension and Plate Read at 60°C, sec | Cycles | Melt Curve<br>Analysis                            |
| Bio-Rad CFX96, CFX384, CFX96<br>Touch, CFX96 Touch Deep Well,<br>CFX384 Touch, CFX Connect | SYBR® only        |  | 2–5                       | 15–30   |        |   |
| Bio-Rad Chromo4, iQ5,<br>MiniOpticon, MyiQ   | Standard          |  | 10–15                     | 15–30   |        | 65–95°C<br>0.5°C<br>increments at<br>2–5 sec/step |
| Applied Biosystems 7500,   | Fast              | 20–30 sec for cDNA                                 | 1–3                       | 20–30   |        |   |
| 7900HT, QuantStudio, StepOne,<br>StepOnePlus, and ViiA 7                                   | Standard          | or<br>2–5 min for genomic DNA                      | 15                        | 60  | 35–40  | (or use   |
| Applied Biosystems 7000 and 7300   | Standard          |  | 15                        | 60  |        | instrument default setting)                       |
| Doobo Limbt Cuplay 490   | Fast              |  | 2–5                       | 15–30   |        | default setting)                                  |
| Roche LightCycler 480  | Standard          |  | 15                        | 60  |        |   |
| QIAGEN Rotor-Gene and Stratagene Mx series   | Fast              |  | 2–5                       | 15–30   |        |   |

#### **Recommendations for Primer Design**

- The iTaq Universal SYBR® Green Supermix and the qPCR cycling protocols have been optimized for assays with a primer melting temperature (T<sub>m</sub>) of 60°C and designed using the open source Primer3 program (http://frodo.wi.mit.edu/) under its default settings. For assays designed using other tools, the primer T<sub>m</sub> should be recalculated using Primer3 for determining annealing/extension temperature
- For best qPCR efficiency, design assays targeting an amplicon size of 70–150 bp. For amplicons >250 bp in length or with high GC or AT content, longer annealing/ extension times (30–60 sec) can be used

## **Quality Control**

iTaq Universal SYBR® Green Supermix demonstrates high PCR efficiency and a wide linear dynamic range. It is manufactured under ISO 13485:2016 to ensure lot-to-lot consistency. This product is free of detectable DNase and RNase activities.

#### **Related Products**

Reverse transcription reagents for real-time qPCR:

- iScript Reverse Transcription Supermix for RT-qPCR (1708840)
- iScript Advanced cDNA Synthesis Kit for RT-qPCR (1725037)
- iScript gDNA Clear cDNA Synthesis Kit (1725034)
- iScript cDNA Synthesis Kit (1708890)

Reagents for real-time qPCR:

- SsoAdvanced Universal SYBR® Green Supermix (1725270)
- iTaq Universal SYBR® Green One-Step Kit (1725150)

PCR primer and probe assays for real-time qPCR:

PrimePCR Assays and Panels

Visit **bio-rad.com/amplification** to learn more about Bio-Rad's complete solution for amplification.

Bio-Rad is a trademark of Bio-Rad Laboratories, Inc. in certain jurisdictions. SYBR is a trademark of Thermo Fisher Scientific Inc. Bio-Rad Laboratories, Inc. is licensed by Thermo Fisher Scientific Inc. to sell reagents containing SYBR Green I for use in real-time PCR, for research purposes only.

All trademarks used herein are the property of their respective owner.