



CERTIFICATE OF COMPLIANCE

LRQA

hereby declares that the certification assessment has demonstrated that

iQ-Check STEC VirX Kit

Manufactured by:
Bio-Rad Laboratories
925 Nobel Drive
Hercules, CA 94547
USA

Supplied by:
Bio-Rad
3 boulevard Raymond Poincaré
92430 Marnes-La-Coquette
France

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report. The summary of the validation report is available on the MicroVal website:

www.microval.org

Reference method: ISO/TS 13136:2012 Microbiology of food and animal feed - Real-Time Polymerase Chain Reaction (PCR) based method for the detection of food-borne pathogens. Horizontal method for the detection of shiga-toxin producing (STEC) and the determination of O26, O103, O111, O145 and O157 serogroups

Scope: a broad range of foods including raw dairy products (up to 25 g test portion), raw meat products (excluding poultry) (up to 375g test portion), flour & raw doughs (up to 375 test portion), multicomponent foods and meal components (up to 25 g test portion), raw produces and fruits (up to 25 g test portion)

The validation and certification has been performed in accordance with EN ISO 16140-2:2016 and the MicroVal Rules and Certification Scheme version 9.1.

Certificate no.: 2021LR96

First approval date: 27 June 2022
Surveillance date: 24 October 2023
Expiry date: 26 June 2026

ISSUED BY:

LRQA Nederland B.V.
Rotterdam, The Netherlands